

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026204.D
 Acq On : 31 Dec 2021 05:57
 Operator : JC/MD
 Sample : PB141758ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141758ZHE#05

Quant Time: Dec 31 21:37:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	246294	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109656	53.647	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.300%
35) Dibromofluoromethane	5.397	113	84588	52.827	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.660%
50) Toluene-d8	8.653	98	296781	49.829	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.660%
62) 4-Bromofluorobenzene	11.085	95	98605	46.506	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.020%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	12087	10.914	ug/l	92
20) Methylene Chloride	2.794	84	9850	4.706	ug/l #	87
43) Isopropyl Acetate	6.348	43	14851	3.043	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
Data File : VX026204.D
Acq On : 31 Dec 2021 05:57
Operator : JC/MD
Sample : PB141758ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141758ZHE#05

Quant Time: Dec 31 21:37:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

