



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123121\
 Data File : VX026155.D
 Acq On : 30 Dec 2021 09:07
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Title : SW846 8260
 Last Update : Thu Dec 30 08:02:21 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.2	7244	PASS
75	95	30	60	56.5	17673	PASS
95	95	100	100	100.0	31256	PASS
96	95	5	9	6.9	2145	PASS
173	174	0.00	2	1.5	336	PASS
174	95	50	100	72.3	22610	PASS
175	174	5	9	7.8	1766	PASS
176	174	95	101	96.1	21721	PASS
177	176	5	9	6.5	1407	PASS