

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044541.D
 Acq On : 30 Dec 2024 14:03
 Operator : JC/MD
 Sample : VX1231MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1231MBL01

Quant Time: Dec 31 00:27:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	105671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	178517	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78181	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	73825	39.846	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	79.700%
35) Dibromofluoromethane	5.379	113	55557	44.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.860%
50) Toluene-d8	8.647	98	213634	51.216	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.440%
62) 4-Bromofluorobenzene	11.079	95	73944	50.732	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044541.D
Acq On : 30 Dec 2024 14:03
Operator : JC/MD
Sample : VX1231MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1231MBL01

Quant Time: Dec 31 00:27:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

