

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044545.D
 Acq On : 30 Dec 2024 15:45
 Operator : JC/MD
 Sample : P5317-18DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20241217DL

Quant Time: Dec 31 00:31:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	200283	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91587	44.273	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.540%
35) Dibromofluoromethane	5.385	113	72531	52.288	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.580%
50) Toluene-d8	8.647	98	257969	55.124	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.240%
62) 4-Bromofluorobenzene	11.079	95	83972	51.351	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.700%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.312	101	2846	1.977	ug/l	91
12) 1,1-Dichloroethene	2.319	96	945	0.715	ug/l #	44
27) cis-1,2-Dichloroethene	4.501	96	750	0.420	ug/l	76
30) Chloroform	5.092	83	1146	0.369	ug/l	96
44) Trichloroethene	7.123	130	49829	36.704	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044545.D
Acq On : 30 Dec 2024 15:45
Operator : JC/MD
Sample : P5317-18DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20241217DL

Quant Time: Dec 31 00:31:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

