

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044547.D
 Acq On : 30 Dec 2024 16:32
 Operator : JC/MD
 Sample : P5376-05 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-07-20241218

Quant Time: Dec 31 00:32:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	120346	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211448	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91910	43.558	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.120%		
35) Dibromofluoromethane	5.379	113	72544	49.536	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.080%		
50) Toluene-d8	8.647	98	261182	52.863	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.720%		
62) 4-Bromofluorobenzene	11.079	95	83314	48.259	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.520%		
Target Compounds						
44) Trichloroethene	7.147	130	654	0.456	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044547.D
Acq On : 30 Dec 2024 16:32
Operator : JC/MD
Sample : P5376-05 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-07-20241218

Quant Time: Dec 31 00:32:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

