

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044548.D
 Acq On : 30 Dec 2024 16:55
 Operator : JC/MD
 Sample : P5376-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB03-20241219

Quant Time: Dec 31 00:32:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	204066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87554	43.354	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.700%
35) Dibromofluoromethane	5.385	113	68991	48.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	8.647	98	246454	51.687	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.380%
62) 4-Bromofluorobenzene	11.079	95	78904	47.358	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.720%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4083	4.959	ug/l	92
20) Methylene Chloride	2.782	84	604	0.391	ug/l #	80
30) Chloroform	5.086	83	23973	7.905	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044548.D
Acq On : 30 Dec 2024 16:55
Operator : JC/MD
Sample : P5376-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB03-20241219

Quant Time: Dec 31 00:32:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

