

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044549.D
 Acq On : 30 Dec 2024 17:18
 Operator : JC/MD
 Sample : P5376-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Dec 31 00:33:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	206326	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	188782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93755	45.396	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.800%
35) Dibromofluoromethane	5.385	113	72737	50.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.800%
50) Toluene-d8	8.647	98	264706	54.907	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.820%
62) 4-Bromofluorobenzene	11.079	95	85597	50.812	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.620%
Target Compounds						
3) Chloromethane	1.295	50	613	0.352	ug/l #	77
16) Acetone	2.380	43	1086	1.290	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044549.D
Acq On : 30 Dec 2024 17:18
Operator : JC/MD
Sample : P5376-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Dec 31 00:33:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

