

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044554.D
 Acq On : 30 Dec 2024 19:14
 Operator : JC/MD
 Sample : P5376-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10-MW01D-20241218

Quant Time: Dec 31 00:36:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	123900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	212555	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	80341	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91926	42.316	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.640%		
35) Dibromofluoromethane	5.379	113	72993	49.583	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.160%		
50) Toluene-d8	8.647	98	258525	52.053	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.100%		
62) 4-Bromofluorobenzene	11.079	95	84006	48.406	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.820%		
Target Compounds						
3) Chloromethane	1.294	50	772	0.421	ug/l #	72
16) Acetone	2.392	43	1143	1.291	ug/l #	37
20) Methylene Chloride	2.776	84	509	0.306	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044554.D
Acq On : 30 Dec 2024 19:14
Operator : JC/MD
Sample : P5376-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10-MW01D-20241218

Quant Time: Dec 31 00:36:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

