

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044557.D
 Acq On : 30 Dec 2024 20:24
 Operator : JC/MD
 Sample : P5376-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20241218

Quant Time: Dec 31 00:37:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200838	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	70367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90558	45.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.420%
35) Dibromofluoromethane	5.385	113	70356	50.580	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	8.647	98	248872	53.033	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.060%
62) 4-Bromofluorobenzene	11.079	95	77596	47.321	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.640%
Target Compounds						
16) Acetone	2.386	43	989	1.211	ug/l	Qvalue # 52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044557.D
Acq On : 30 Dec 2024 20:24
Operator : JC/MD
Sample : P5376-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20241218

Quant Time: Dec 31 00:37:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

