

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044559.D
 Acq On : 30 Dec 2024 21:11
 Operator : JC/MD
 Sample : P5376-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT162S1-20241219

Quant Time: Dec 31 00:38:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	114172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	199344	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74318	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90224	45.071	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.140%		
35) Dibromofluoromethane	5.379	113	68868	49.881	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.760%		
50) Toluene-d8	8.647	98	245495	52.705	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.420%		
62) 4-Bromofluorobenzene	11.079	95	79780	49.018	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.040%		
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.306	96	459	0.359	ug/l #	59
16) Acetone	2.386	43	1083	1.327	ug/l #	73
20) Methylene Chloride	2.788	84	671	0.438	ug/l #	47
24) 1,1-Dichloroethane	3.605	63	3543	1.324	ug/l #	94
30) Chloroform	5.092	83	1390	0.462	ug/l #	73
44) Trichloroethene	7.123	130	16730	12.381	ug/l	100
64) Tetrachloroethene	9.281	164	784	0.675	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044559.D
Acq On : 30 Dec 2024 21:11
Operator : JC/MD
Sample : P5376-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT162S1-20241219

Quant Time: Dec 31 00:38:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

