

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044562.D
 Acq On : 30 Dec 2024 22:21
 Operator : JC/MD
 Sample : P5376-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT119S1-20241219

Quant Time: Dec 31 00:40:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	119469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211589	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94325	45.031	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.060%		
35) Dibromofluoromethane	5.385	113	73762	50.334	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.660%		
50) Toluene-d8	8.647	98	260390	52.668	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.340%		
62) 4-Bromofluorobenzene	11.079	95	83857	48.541	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.080%		
Target Compounds						
16) Acetone	2.380	43	969	1.135	ug/l	90
44) Trichloroethene	7.129	130	2900	2.022	ug/l	62
64) Tetrachloroethene	9.275	164	620	0.503	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044562.D
Acq On : 30 Dec 2024 22:21
Operator : JC/MD
Sample : P5376-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT119S1-20241219

Quant Time: Dec 31 00:40:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

