

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044563.D
 Acq On : 30 Dec 2024 22:44
 Operator : JC/MD
 Sample : P5376-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10-MW01S-20241218

Quant Time: Dec 31 00:40:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	192338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85418	44.590	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.180%		
35) Dibromofluoromethane	5.385	113	66327	49.791	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.580%		
50) Toluene-d8	8.647	98	240413	53.494	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	106.980%		
62) 4-Bromofluorobenzene	11.079	95	75918	48.344	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.680%		
Target Compounds						
16) Acetone	2.392	43	1032	1.321	ug/l	94
19) Methyl tert-butyl Ether	3.111	73	5220	1.092	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	1080	0.653	ug/l	96
44) Trichloroethene	7.135	130	985	0.756	ug/l	87
64) Tetrachloroethene	9.275	164	5610	4.896	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044563.D
Acq On : 30 Dec 2024 22:44
Operator : JC/MD
Sample : P5376-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10-MW01S-20241218

Quant Time: Dec 31 00:40:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

