

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044569.D
 Acq On : 31 Dec 2024 10:26
 Operator : JC/MD
 Sample : VX1231MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1231MBL02

Quant Time: Jan 01 01:36:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	125340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	215197	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	194646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	85873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89308	40.638	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	81.280%	
35) Dibromofluoromethane	5.379	113	71688	48.099	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.200%	
50) Toluene-d8	8.647	98	257495	51.209	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.420%	
62) 4-Bromofluorobenzene	11.079	95	86974	49.501	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	99.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044569.D
Acq On : 31 Dec 2024 10:26
Operator : JC/MD
Sample : VX1231MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1231MBL02

Quant Time: Jan 01 01:36:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

