

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044570.D
 Acq On : 31 Dec 2024 10:49
 Operator : JC/MD
 Sample : VX1231WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1231WBL02

Quant Time: Jan 01 01:36:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	128703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	217401	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	189142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	80019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95381	42.268	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.540%
35) Dibromofluoromethane	5.379	113	75909	50.414	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.820%
50) Toluene-d8	8.647	98	268209	52.799	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.600%
62) 4-Bromofluorobenzene	11.079	95	84593	47.658	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.320%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044570.D
Acq On : 31 Dec 2024 10:49
Operator : JC/MD
Sample : VX1231WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1231WBL02

Quant Time: Jan 01 01:36:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

