

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
 Data File : VX044573.D
 Acq On : 31 Dec 2024 12:14
 Operator : JC/MD
 Sample : P5376-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-07-20241218

Quant Time: Jan 01 01:39:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	126196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	215542	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94227	42.586	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.180%
35) Dibromofluoromethane	5.385	113	74374	49.821	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	8.647	98	266792	52.973	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.940%
62) 4-Bromofluorobenzene	11.079	95	87911	49.954	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1493	1.655	ug/l #	83
19) Methyl tert-butyl Ether	3.117	73	41720	7.558	ug/l	100
44) Trichloroethene	7.123	130	6087	4.166	ug/l	95
64) Tetrachloroethene	9.275	164	1489	1.176	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX123124\
Data File : VX044573.D
Acq On : 31 Dec 2024 12:14
Operator : JC/MD
Sample : P5376-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-07-20241218

Quant Time: Jan 01 01:39:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

