

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046339.D
 Acq On : 23 May 2025 12:56
 Operator : JC/MD
 Sample : Q2120-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-EB-20250522

Quant Time: May 23 23:00:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132414	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65571	52.815	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.620%	
35) Dibromofluoromethane	5.385	113	47284	49.589	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.180%	
50) Toluene-d8	8.647	98	163179	49.444	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.880%	
62) 4-Bromofluorobenzene	11.079	95	62331	49.237	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.480%	
Target Compounds						
16) Acetone	2.380	43	7404	14.874	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046339.D
Acq On : 23 May 2025 12:56
Operator : JC/MD
Sample : Q2120-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-EB-20250522

Quant Time: May 23 23:00:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

