

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046341.D
 Acq On : 23 May 2025 13:43
 Operator : JC/MD
 Sample : Q2118-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-460-462

Quant Time: May 23 23:01:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66245	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133836	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66015	53.453	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.900%
35) Dibromofluoromethane	5.385	113	48522	50.347	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.700%
50) Toluene-d8	8.647	98	167542	50.227	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.460%
62) 4-Bromofluorobenzene	11.079	95	66411	51.903	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.800%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1448	2.924	ug/l #	83
17) Carbon Disulfide	2.508	76	844	0.453	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046341.D
Acq On : 23 May 2025 13:43
Operator : JC/MD
Sample : Q2118-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-460-462

Quant Time: May 23 23:01:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

