

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046342.D
 Acq On : 23 May 2025 14:06
 Operator : JC/MD
 Sample : Q2118-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-420-422

Quant Time: May 23 23:02:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65112	52.689	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.380%
35) Dibromofluoromethane	5.379	113	46984	50.458	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.920%
50) Toluene-d8	8.647	98	163256	50.656	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.320%
62) 4-Bromofluorobenzene	11.079	95	61565	49.800	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.600%
Target Compounds						
16) Acetone	2.386	43	1348	2.721	ug/l #	85
17) Carbon Disulfide	2.508	76	667	0.358	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046342.D
Acq On : 23 May 2025 14:06
Operator : JC/MD
Sample : Q2118-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-420-422

Quant Time: May 23 23:02:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

