

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046343.D
 Acq On : 23 May 2025 14:29
 Operator : JC/MD
 Sample : Q2118-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-450-452

Quant Time: May 23 23:02:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122489	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62133	54.027	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.060%		
35) Dibromofluoromethane	5.385	113	45376	51.444	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.880%		
50) Toluene-d8	8.647	98	156196	51.163	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.320%		
62) 4-Bromofluorobenzene	11.079	95	58715	50.139	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.280%		
Target Compounds						
16) Acetone	2.380	43	1246	2.702	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046343.D
Acq On : 23 May 2025 14:29
Operator : JC/MD
Sample : Q2118-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-450-452

Quant Time: May 23 23:02:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

