

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046344.D
 Acq On : 23 May 2025 14:53
 Operator : JC/MD
 Sample : Q2118-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-DUP-20250520

Quant Time: May 23 23:02:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63254	54.100	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.200%	
35) Dibromofluoromethane	5.379	113	45684	50.859	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.720%	
50) Toluene-d8	8.647	98	156867	50.456	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.920%	
62) 4-Bromofluorobenzene	11.079	95	59241	49.676	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.360%	
Target Compounds						
16) Acetone	2.386	43	1410	3.008	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046344.D
Acq On : 23 May 2025 14:53
Operator : JC/MD
Sample : Q2118-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250520

Quant Time: May 23 23:02:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

