

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046346.D
 Acq On : 23 May 2025 15:39
 Operator : JC/MD
 Sample : Q2118-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-520-522

Quant Time: May 23 23:03:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134921	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66856	53.253	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.500%
35) Dibromofluoromethane	5.379	113	49552	51.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.000%
50) Toluene-d8	8.647	98	169340	50.358	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.720%
62) 4-Bromofluorobenzene	11.079	95	64477	49.986	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.980%
Target Compounds						
3) Chloromethane	1.307	50	785	0.785	ug/l #	85
16) Acetone	2.380	43	1203	2.390	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046346.D
Acq On : 23 May 2025 15:39
Operator : JC/MD
Sample : Q2118-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-520-522

Quant Time: May 23 23:03:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

