

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
 Data File : VX046349.D
 Acq On : 23 May 2025 16:49
 Operator : JC/MD
 Sample : Q2114-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: May 23 23:04:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49600	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62190	53.920	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.840%
35) Dibromofluoromethane	5.379	113	45440	51.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.380%
50) Toluene-d8	8.647	98	155329	50.554	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.100%
62) 4-Bromofluorobenzene	11.079	95	59792	50.732	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052325\
Data File : VX046349.D
Acq On : 23 May 2025 16:49
Operator : JC/MD
Sample : Q2114-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: May 23 23:04:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

