

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
 Data File : VX029543.D
 Acq On : 17 Jun 2022 10:04
 Operator : JC/MD
 Sample : VX0617WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0617WBL01

Quant Time: Jun 17 10:41:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	161142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	270893	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157426	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158133	64.993	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	129.980%
35) Dibromofluoromethane	5.391	113	134504	58.405	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	116.800%
50) Toluene-d8	8.653	98	480601	59.006	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	118.020%
62) 4-Bromofluorobenzene	11.079	95	180867	56.744	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
Data File : VX029543.D
Acq On : 17 Jun 2022 10:04
Operator : JC/MD
Sample : VX0617WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0617WBL01

Quant Time: Jun 17 10:41:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

