

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
 Data File : VX029544.D
 Acq On : 17 Jun 2022 12:51
 Operator : JC/MD
 Sample : VX0617WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0617WBL02

Quant Time: Jun 17 13:25:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	274842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	463704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181413	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	243384	58.471	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.940%
35) Dibromofluoromethane	5.385	113	208978	53.011	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.020%
50) Toluene-d8	8.647	98	784471	56.266	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	112.540%
62) 4-Bromofluorobenzene	11.079	95	292280	53.569	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
Data File : VX029544.D
Acq On : 17 Jun 2022 12:51
Operator : JC/MD
Sample : VX0617WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0617WBL02

Quant Time: Jun 17 13:25:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

