

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
 Data File : VX029547.D
 Acq On : 17 Jun 2022 17:43
 Operator : JC/MD
 Sample : N3314-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ER-HEATER-COMP

Quant Time: Jun 18 00:47:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	169369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	277940	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	282994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132700	51.524	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.040%
35) Dibromofluoromethane	5.391	113	111306	47.106	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.220%
50) Toluene-d8	8.653	98	409494	49.001	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.085	95	157544	48.174	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.340%
Target Compounds						
44) Trichloroethene	7.129	130	2554	0.807	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
Data File : VX029547.D
Acq On : 17 Jun 2022 17:43
Operator : JC/MD
Sample : N3314-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ER-HEATER-COMP

Quant Time: Jun 18 00:47:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

