

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
 Data File : VX029551.D
 Acq On : 17 Jun 2022 19:17
 Operator : JC/MD
 Sample : N3334-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-1A-8

Quant Time: Jun 18 00:47:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	160383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	267004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	275292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	120246	49.226	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.460%		
35) Dibromofluoromethane	5.391	113	103568	45.626	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.260%		
50) Toluene-d8	8.653	98	363523	45.282	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	90.560%		
62) 4-Bromofluorobenzene	11.085	95	147971	47.100	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.200%		
Target Compounds					Qvalue	
16) Acetone	2.379	43	64481	49.698	ug/l	94
20) Methylene Chloride	2.788	84	56627	19.544	ug/l	96
25) 2-Butanone	4.568	43	29128	14.192	ug/l	93
31) Cyclohexane	5.483	56	5488	1.262	ug/l	86
39) Methylcyclohexane	7.385	83	7998	1.732	ug/l	92
43) Isopropyl Acetate	6.348	43	27294	4.694	ug/l	96
44) Trichloroethene	7.122	130	4502	1.481	ug/l	85
52) Toluene	8.726	92	7898	1.102	ug/l	91
67) Ethyl Benzene	10.195	91	52203	3.914	ug/l	99
68) m/p-Xylenes	10.305	106	35360	6.852	ug/l	98
69) o-Xylene	10.646	106	101398	20.222	ug/l	100
73) Isopropylbenzene	10.963	105	39748	2.926	ug/l	99
78) n-propylbenzene	11.305	91	43455	2.819	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	159614	13.835	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	356268	31.287	ug/l	98
85) sec-Butylbenzene	11.890	105	17946	1.306	ug/l	83
86) p-Isopropyltoluene	12.012	119	19446	1.687	ug/l	94
95) Naphthalene	13.780	128	240951	19.160	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061722\
Data File : VX029551.D
Acq On : 17 Jun 2022 19:17
Operator : JC/MD
Sample : N3334-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-1A-8

Quant Time: Jun 18 00:47:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

