

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030820.D
 Acq On : 26 Aug 2022 13:29
 Operator : JC/MD
 Sample : N4245-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT161S1-20220817

Quant Time: Aug 27 02:38:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	227500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	389090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	339870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	143180	53.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.020%	
35) Dibromofluoromethane	5.391	113	121227	50.695	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.400%	
50) Toluene-d8	8.653	98	447603	53.716	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.440%	
62) 4-Bromofluorobenzene	11.085	95	146639	47.268	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.540%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1532	0.743	ug/l	95
19) Methyl tert-butyl Ether	3.123	73	3711	0.520	ug/l	90
24) 1,1-Dichloroethane	3.611	63	13567	3.123	ug/l	96
30) Chloroform	5.105	83	1920	0.420	ug/l #	63
44) Trichloroethene	7.129	130	2391	0.849	ug/l	81
52) Toluene	8.726	92	1600	0.263	ug/l	91
64) Tetrachloroethene	9.281	164	1032	0.385	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030820.D
Acq On : 26 Aug 2022 13:29
Operator : JC/MD
Sample : N4245-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT161S1-20220817

Quant Time: Aug 27 02:38:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

