

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030823.D
 Acq On : 26 Aug 2022 14:40
 Operator : JC/MD
 Sample : N4245-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D1-20220816DL

Quant Time: Aug 27 02:39:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	222584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	383834	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	140313	53.592	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.180%	
35) Dibromofluoromethane	5.391	113	119277	50.563	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.120%	
50) Toluene-d8	8.653	98	447798	54.475	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.940%	
62) 4-Bromofluorobenzene	11.085	95	148400	48.490	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.980%	
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	9095	3.397	ug/l	91
44) Trichloroethene	7.129	130	67648	24.364	ug/l	97
64) Tetrachloroethene	9.275	164	4991	1.862	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030823.D
Acq On : 26 Aug 2022 14:40
Operator : JC/MD
Sample : N4245-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D1-20220816DL

Quant Time: Aug 27 02:39:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

