

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030826.D
 Acq On : 26 Aug 2022 15:51
 Operator : JC/MD
 Sample : N4245-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT163S1-20220817

Quant Time: Aug 27 02:40:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	216475	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.775	114	365619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	320291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	138727	54.481	ug/l	0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.960%	
35) Dibromofluoromethane	5.397	113	114847	51.111	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.220%	
50) Toluene-d8	8.653	98	426692	54.493	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.980%	
62) 4-Bromofluorobenzene	11.085	95	138620	47.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.100%	
Target Compounds						
16) Acetone	2.386	43	2663	2.263	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030826.D
Acq On : 26 Aug 2022 15:51
Operator : JC/MD
Sample : N4245-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT163S1-20220817

Quant Time: Aug 27 02:40:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

