

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030829.D
 Acq On : 26 Aug 2022 17:02
 Operator : JC/MD
 Sample : N4262-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102S

Quant Time: Aug 27 02:40:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	206153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	345274	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	303733	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	132219	54.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.060%	
35) Dibromofluoromethane	5.391	113	110382	52.018	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.040%	
50) Toluene-d8	8.653	98	407136	55.060	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.120%	
62) 4-Bromofluorobenzene	11.085	95	130019	47.229	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.460%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.313	96	456	0.244	ug/l #	87
27) cis-1,2-Dichloroethene	4.507	96	623	0.251	ug/l #	17
44) Trichloroethene	7.129	130	41625	16.666	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030829.D
Acq On : 26 Aug 2022 17:02
Operator : JC/MD
Sample : N4262-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102S

Quant Time: Aug 27 02:40:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

