

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030830.D
 Acq On : 26 Aug 2022 17:25
 Operator : JC/MD
 Sample : N4262-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102I

Quant Time: Aug 27 02:40:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	204858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	346545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	314349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	130129	54.003	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.000%	
35) Dibromofluoromethane	5.391	113	111884	52.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.060%	
50) Toluene-d8	8.652	98	412174	55.537	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.080%	
62) 4-Bromofluorobenzene	11.085	95	137930	49.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.840%	
Target Compounds						
					Qvalue	
21) trans-1,2-Dichloroethene	3.093	96	1618	0.811	ug/l #	75
27) cis-1,2-Dichloroethene	4.501	96	8039	3.262	ug/l	91
30) Chloroform	5.098	83	2881	0.700	ug/l #	80
44) Trichloroethene	7.134	130	51939	20.719	ug/l	99
64) Tetrachloroethene	9.274	164	87896	35.496	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030830.D
Acq On : 26 Aug 2022 17:25
Operator : JC/MD
Sample : N4262-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102I

Quant Time: Aug 27 02:40:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

