

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030832.D
 Acq On : 26 Aug 2022 18:13
 Operator : JC/MD
 Sample : N4262-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-101I

Quant Time: Aug 27 02:41:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	215868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	362261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	327474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	137429	54.123	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.240%			
35) Dibromofluoromethane	5.397	113	115176	51.732	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.460%			
50) Toluene-d8	8.652	98	429091	55.308	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.620%			
62) 4-Bromofluorobenzene	11.085	95	140985	48.811	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.620%			
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	2.318	96	37037	18.927	ug/l	92
24) 1,1-Dichloroethane	3.617	63	77547	18.812	ug/l	98
27) cis-1,2-Dichloroethene	4.501	96	54716	21.071	ug/l	97
30) Chloroform	5.110	83	2567	0.592	ug/l	98
32) 1,1,1-Trichloroethane	5.403	97	6309	1.799	ug/l	99
44) Trichloroethene	7.128	130	52651	20.092	ug/l	96
49) 1,4-Dioxane	7.665	88	4982	74.849	ug/l	98
64) Tetrachloroethene	9.280	164	17973	6.967	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030832.D
Acq On : 26 Aug 2022 18:13
Operator : JC/MD
Sample : N4262-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-101I

Quant Time: Aug 27 02:41:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

