

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030834.D
 Acq On : 26 Aug 2022 19:00
 Operator : JC/MD
 Sample : N4262-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-107(42.5FT)

Quant Time: Aug 27 02:41:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	204280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	343587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	305258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	130241	54.202	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.400%			
35) Dibromofluoromethane	5.397	113	109755	51.976	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.960%			
50) Toluene-d8	8.653	98	410573	55.797	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.600%			
62) 4-Bromofluorobenzene	11.085	95	131858	48.132	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.260%			
Target Compounds						
21) trans-1,2-Dichloroethene	3.099	96	1106	0.556	ug/l #	77
27) cis-1,2-Dichloroethene	4.495	96	6730	2.739	ug/l	94
30) Chloroform	5.105	83	2996	0.730	ug/l	91
44) Trichloroethene	7.135	130	45247	18.205	ug/l	98
64) Tetrachloroethene	9.275	164	81007	33.688	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030834.D
Acq On : 26 Aug 2022 19:00
Operator : JC/MD
Sample : N4262-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-107(42.5FT)

Quant Time: Aug 27 02:41:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

