

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030841.D
 Acq On : 27 Aug 2022 01:42
 Operator : JC/MD
 Sample : VX0826WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0826WBL02

Quant Time: Aug 27 06:28:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	219835	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	371181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	336204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146861	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142592	55.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.280%	
35) Dibromofluoromethane	5.385	113	119671	52.459	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.920%	
50) Toluene-d8	8.647	98	438998	55.225	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.440%	
62) 4-Bromofluorobenzene	11.079	95	148677	50.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030841.D
Acq On : 27 Aug 2022 01:42
Operator : JC/MD
Sample : VX0826WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0826WBL02

Quant Time: Aug 27 06:28:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

