

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030850.D
 Acq On : 27 Aug 2022 06:16
 Operator : JC/MD
 Sample : N4281-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20220818

Quant Time: Aug 27 06:47:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	227930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	394099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	349630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	150110	55.989	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.980%	
35) Dibromofluoromethane	5.391	113	122174	50.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.880%	
50) Toluene-d8	8.653	98	462959	54.852	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.700%	
62) 4-Bromofluorobenzene	11.085	95	153602	48.883	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.760%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	2891	1.369	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.331	101	36346	16.375	ug/l	97
12) 1,1-Dichloroethene	2.319	96	14789	7.158	ug/l	91
24) 1,1-Dichloroethane	3.617	63	4064	0.934	ug/l #	90
27) cis-1,2-Dichloroethene	4.501	96	6667	2.432	ug/l	94
30) Chloroform	5.105	83	8526	1.861	ug/l	91
32) 1,1,1-Trichloroethane	5.385	97	1912	0.516	ug/l #	46
38) Carbon Tetrachloride	5.684	117	8040	2.311	ug/l	91
42) 1,2-Dichloroethane	6.092	62	1156	0.330	ug/l	92
44) Trichloroethene	7.129	130	799204	280.340	ug/l	88
55) 1,1,2-Trichloroethane	9.159	97	2184	0.798	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030850.D
Acq On : 27 Aug 2022 06:16
Operator : JC/MD
Sample : N4281-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20220818

Quant Time: Aug 27 06:47:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

