

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030852.D
 Acq On : 27 Aug 2022 07:03
 Operator : JC/MD
 Sample : N4281-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D3-20220818

Quant Time: Aug 27 22:20:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	222616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	383378	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	338527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	145249	55.469	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.940%	
35) Dibromofluoromethane	5.391	113	121261	51.465	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.940%	
50) Toluene-d8	8.653	98	449067	54.694	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.380%	
62) 4-Bromofluorobenzene	11.085	95	146523	47.934	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.860%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	1662	0.806	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.331	101	8531	3.935	ug/l	92
12) 1,1-Dichloroethene	2.325	96	1044	0.517	ug/l #	90
27) cis-1,2-Dichloroethene	4.495	96	2289	0.855	ug/l	84
44) Trichloroethene	7.129	130	178916	64.514	ug/l	99
64) Tetrachloroethene	9.275	164	4413	1.655	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030852.D
Acq On : 27 Aug 2022 07:03
Operator : JC/MD
Sample : N4281-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D3-20220818

Quant Time: Aug 27 22:20:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

