

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030853.D
 Acq On : 27 Aug 2022 07:26
 Operator : JC/MD
 Sample : N4281-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20220818

Quant Time: Aug 27 22:21:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	221889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	382397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	347682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149649	57.337	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.680%			
35) Dibromofluoromethane	5.391	113	120525	51.284	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.560%			
50) Toluene-d8	8.653	98	457083	55.813	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.620%			
62) 4-Bromofluorobenzene	11.085	95	152123	49.894	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.780%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	809	0.393	ug/l	75
7) Trichlorofluoromethane	1.892	101	1008	0.297	ug/l #	77
9) 1,1,2-Trichlorotrifluo...	2.331	101	10669	4.937	ug/l	97
12) 1,1-Dichloroethene	2.318	96	2042	1.015	ug/l	90
27) cis-1,2-Dichloroethene	4.495	96	5165	1.935	ug/l	90
38) Carbon Tetrachloride	5.696	117	1223	0.362	ug/l #	80
44) Trichloroethene	7.129	130	561402	202.952	ug/l	95
64) Tetrachloroethene	9.274	164	13571	4.955	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030853.D
Acq On : 27 Aug 2022 07:26
Operator : JC/MD
Sample : N4281-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20220818

Quant Time: Aug 27 22:21:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

