

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X082622\
 Data File : VX030840.D
 Acq On : 27 Aug 2022 00:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 27 06:28:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	50.000	43.487	13.0	70	0.00
3 P	Chloromethane	50.000	52.395	-4.8	81	0.00
4 C	Vinyl Chloride	50.000	47.175	5.7#	73	0.00
5 T	Bromomethane	50.000	54.177	-8.4	95	0.00
6 T	Chloroethane	50.000	55.223	-10.4	84	0.00
7 T	Trichlorofluoromethane	50.000	46.254	7.5	78	0.00
8 T	Diethyl Ether	50.000	59.089	-18.2	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	43.330	13.3	72	0.00
10 T	Methyl Iodide	50.000	45.190	9.6	73	0.00
11 T	Tert butyl alcohol	250.000	302.822	-21.1	105	0.00
12 CM	1,1-Dichloroethene	50.000	43.432	13.1#	74	0.00
13 T	Acrolein	250.000	327.412	-31.0#	106	0.00
14 T	Allyl chloride	50.000	49.304	1.4	83	0.00
15 T	Acrylonitrile	250.000	306.835	-22.7	99	0.00
16 T	Acetone	250.000	299.188	-19.7	98	0.00
17 T	Carbon Disulfide	50.000	43.515	13.0	68	0.00
18 T	Methyl Acetate	50.000	61.241	-22.5	100	0.00
19 T	Methyl tert-butyl Ether	50.000	59.485	-19.0	95	0.00
20 T	Methylene Chloride	50.000	58.145	-16.3	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.793	8.4	78	0.00
22 T	Diisopropyl ether	50.000	58.458	-16.9	95	0.00
23 T	Vinyl Acetate	250.000	303.672	-21.5	96	0.00
24 P	1,1-Dichloroethane	50.000	51.808	-3.6	86	0.00
25 T	2-Butanone	250.000	307.123	-22.8	100	0.00
26 T	2,2-Dichloropropane	50.000	31.916	36.2#	50	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.741	-1.5	85	0.00
28 T	Bromochloromethane	50.000	57.376	-14.8	94	0.00
29 T	Tetrahydrofuran	250.000	307.007	-22.8	100	0.00
30 C	Chloroform	50.000	53.617	-7.2#	91	0.00
31 T	Cyclohexane	50.000	44.744	10.5	72	0.00
32 T	1,1,1-Trichloroethane	50.000	49.948	0.1	80	0.00
33 S	1,2-Dichloroethane-d4	50.000	55.627	-11.3	96	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	82	0.00
35 S	Dibromofluoromethane	50.000	50.492	-1.0	89	0.00
36 T	1,1-Dichloropropene	50.000	42.964	14.1	78	0.00
37 T	Ethyl Acetate	50.000	56.579	-13.2	100	0.00
38 T	Carbon Tetrachloride	50.000	42.944	14.1	76	0.00
39 T	Methylcyclohexane	50.000	39.625	20.8	70	0.00
40 TM	Benzene	50.000	48.703	2.6	87	0.00
41 T	Methacrylonitrile	50.000	58.307	-16.6	100	0.00
42 TM	1,2-Dichloroethane	50.000	55.046	-10.1	98	-0.01
43 T	Isopropyl Acetate	50.000	57.319	-14.6	98	0.00
44 TM	Trichloroethene	50.000	44.570	10.9	79	0.00
45 C	1,2-Dichloropropane	50.000	50.906	-1.8#	88	0.00
46 T	Dibromomethane	50.000	51.355	-2.7	91	0.00
47 T	Bromodichloromethane	50.000	53.528	-7.1	91	0.00
48 T	Methyl methacrylate	50.000	58.011	-16.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030840.D
 Acq On : 27 Aug 2022 00:07
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 27 06:28:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 26 02:03:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1175.301	-17.5	104	0.00
50 S	Toluene-d8	50.000	49.515	1.0	87	0.00
51 T	4-Methyl-2-Pentanone	250.000	300.897	-20.4	102	0.00
52 CM	Toluene	50.000	49.637	0.7#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	52.842	-5.7	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.697	-1.4	84	0.00
55 T	1,1,2-Trichloroethane	50.000	55.354	-10.7	95	0.00
56 T	Ethyl methacrylate	50.000	58.518	-17.0	95	0.00
57 T	1,3-Dichloropropane	50.000	55.154	-10.3	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	302.805	-21.1	97	0.00
59 T	2-Hexanone	250.000	307.405	-23.0	102	0.00
60 T	Dibromochloromethane	50.000	55.187	-10.4	92	0.00
61 T	1,2-Dibromoethane	50.000	55.372	-10.7	94	0.00
62 S	4-Bromofluorobenzene	50.000	52.028	-4.1	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	88	0.00
64 T	Tetrachloroethene	50.000	42.814	14.4	80	0.00
65 PM	Chlorobenzene	50.000	46.641	6.7	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.867	2.3	88	0.00
67 C	Ethyl Benzene	50.000	46.068	7.9#	84	0.00
68 T	m/p-Xylenes	100.000	93.203	6.8	85	0.00
69 T	o-Xylene	50.000	47.767	4.5	87	0.00
70 T	Styrene	50.000	50.440	-0.9	88	0.00
71 P	Bromoform	50.000	49.674	0.7	89	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	84	0.00
73 T	Isopropylbenzene	50.000	46.304	7.4	83	0.00
74 T	N-ethyl acetate	50.000	58.018	-16.0	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	54.121	-8.2	100	0.00
76 T	1,2,3-Trichloropropane	50.000	53.140	-6.3	86	0.00
77 T	Bromobenzene	50.000	48.287	3.4	89	0.00
78 T	n-propylbenzene	50.000	47.126	5.7	83	0.00
79 T	2-Chlorotoluene	50.000	46.944	6.1	86	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.096	5.8	84	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.825	0.3	84	0.00
82 T	4-Chlorotoluene	50.000	48.352	3.3	86	0.00
83 T	tert-Butylbenzene	50.000	47.121	5.8	81	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.672	0.7	85	0.00
85 T	sec-Butylbenzene	50.000	47.233	5.5	80	0.00
86 T	p-Isopropyltoluene	50.000	47.778	4.4	81	0.00
87 T	1,3-Dichlorobenzene	50.000	48.105	3.8	80	0.00
88 T	1,4-Dichlorobenzene	50.000	47.681	4.6	83	0.00
89 T	n-Butylbenzene	50.000	45.280	9.4	73	0.00
90 T	Hexachloroethane	50.000	43.021	14.0	74	0.00
91 T	1,2-Dichlorobenzene	50.000	47.283	5.4	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.350	1.3	85	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.743	6.5	77	0.00
94 T	Hexachlorobutadiene	50.000	41.471	17.1	70	0.00
95 T	Naphthalene	50.000	52.351	-4.7	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
Data File : VX030840.D
Acq On : 27 Aug 2022 00:07
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 27 06:28:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 26 02:03:28 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	46.376	7.2	75	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6