



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082622\
 Data File : VX030839.D
 Acq On : 26 Aug 2022 23:25
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082522W.M
 Title : SW846 8260
 Last Update : Fri Aug 26 02:03:28 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	10764	PASS
75	95	30	60	50.2	28733	PASS
95	95	100	100	100.0	57264	PASS
96	95	5	9	6.2	3576	PASS
173	174	0.00	2	1.4	647	PASS
174	95	50	100	80.2	45907	PASS
175	174	5	9	7.1	3264	PASS
176	174	95	101	95.0	43592	PASS
177	176	5	9	6.9	2990	PASS