

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033279.D
 Acq On : 07 Dec 2022 12:05
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Dec 07 12:45:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	150997	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97255	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	105285	159.756	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	319.520%#
35) Dibromofluoromethane	5.385	113	85773	126.294	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	252.580%#
50) Toluene-d8	8.647	98	305810	215.245	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	430.480%#
62) 4-Bromofluorobenzene	11.079	95	105535	97.148	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	194.300%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033279.D
Acq On : 07 Dec 2022 12:05
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Dec 07 12:45:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

