

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033285.D
 Acq On : 07 Dec 2022 16:45
 Operator : JC/MD
 Sample : VX1207WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1207WBL02

Quant Time: Dec 08 01:42:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95432	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47544	48.952	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.900%
35) Dibromofluoromethane	5.385	113	52410	49.116	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	8.647	98	110673	49.580	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.160%
62) 4-Bromofluorobenzene	11.079	95	82921	48.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033285.D
Acq On : 07 Dec 2022 16:45
Operator : JC/MD
Sample : VX1207WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1207WBL02

Quant Time: Dec 08 01:42:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

