

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033288.D
 Acq On : 07 Dec 2022 17:59
 Operator : JC/MD
 Sample : N5865-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-5SR

Quant Time: Dec 08 01:43:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201834	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37552	46.434	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.860%
35) Dibromofluoromethane	5.385	113	40311	44.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.800%
50) Toluene-d8	8.647	98	84987	44.751	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.500%#
62) 4-Bromofluorobenzene	11.079	95	62760	43.221	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.440%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	808	0.695	ug/l #	70
17) Carbon Disulfide	2.514	76	1148	0.391	ug/l #	94
27) cis-1,2-Dichloroethene	4.495	96	1801	1.205	ug/l	90
38) Carbon Tetrachloride	5.550	117	12926	6.518	ug/l #	15

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033288.D
Acq On : 07 Dec 2022 17:59
Operator : JC/MD
Sample : N5865-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-5SR

Quant Time: Dec 08 01:43:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

