

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033289.D
 Acq On : 07 Dec 2022 18:22
 Operator : JC/MD
 Sample : N5865-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-5D

Quant Time: Dec 08 01:43:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223216	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	44160	49.904	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.800%
35) Dibromofluoromethane	5.391	113	47866	47.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.360%
50) Toluene-d8	8.653	98	100531	47.866	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.740%
62) 4-Bromofluorobenzene	11.079	95	75724	47.154	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.300%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	1032	0.631	ug/l	78
44) Trichloroethene	7.135	130	403	0.256	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033289.D
Acq On : 07 Dec 2022 18:22
Operator : JC/MD
Sample : N5865-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-5D

Quant Time: Dec 08 01:43:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

