

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033293.D
 Acq On : 07 Dec 2022 19:55
 Operator : JC/MD
 Sample : N5865-11DL 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-01DL

Quant Time: Dec 08 01:44:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216485	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85098	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46119	54.531	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.060%
35) Dibromofluoromethane	5.391	113	48670	49.984	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.960%
50) Toluene-d8	8.653	98	97403	47.818	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.640%
62) 4-Bromofluorobenzene	11.079	95	72614	46.623	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.240%
Target Compounds						
27) cis-1,2-Dichloroethene	4.507	96	661	0.423	ug/l	87
44) Trichloroethene	7.129	130	10726	7.028	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033293.D
Acq On : 07 Dec 2022 19:55
Operator : JC/MD
Sample : N5865-11DL 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-01DL

Quant Time: Dec 08 01:44:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

