

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033294.D
 Acq On : 07 Dec 2022 20:18
 Operator : JC/MD
 Sample : N5865-01DL 400X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-7IDL

Quant Time: Dec 08 01:44:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142257	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228846	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202713	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45913	50.309	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.620%	
35) Dibromofluoromethane	5.391	113	49387	47.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.960%	
50) Toluene-d8	8.647	98	103766	48.190	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.380%	
62) 4-Bromofluorobenzene	11.079	95	76310	46.349	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.700%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	518	0.402	ug/l	87
21) trans-1,2-Dichloroethene	3.093	96	845	0.599	ug/l	90
27) cis-1,2-Dichloroethene	4.489	96	430050	255.066	ug/l	89
44) Trichloroethene	7.129	130	142943	88.607	ug/l	99
68) m/p-Xylenes	10.305	106	1104	0.397	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033294.D
Acq On : 07 Dec 2022 20:18
Operator : JC/MD
Sample : N5865-01DL 400X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-7IDL

Quant Time: Dec 08 01:44:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

