

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033295.D
 Acq On : 07 Dec 2022 20:41
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Dec 08 01:44:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197501	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45459	62.441	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 124.880%#	
35) Dibromofluoromethane	5.391	113	47177	53.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.220%	
50) Toluene-d8	8.653	98	90937	48.935	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.880%	
62) 4-Bromofluorobenzene	11.079	95	66826	47.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.060%	
Target Compounds						Qvalue
20) Methylene Chloride	2.788	84	605	0.504	ug/l	89
27) cis-1,2-Dichloroethene	4.495	96	707	0.526	ug/l	79
44) Trichloroethene	7.135	130	371	0.266	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033295.D
Acq On : 07 Dec 2022 20:41
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Dec 08 01:44:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

