

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033297.D
 Acq On : 07 Dec 2022 21:28
 Operator : JC/MD
 Sample : N5911-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A3594TOG

Quant Time: Dec 08 01:44:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/08/2022
 Supervised By : Mahesh Dadoda 12/08/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240716	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47179	48.541	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.080%		
35) Dibromofluoromethane	5.385	113	52513	48.502	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.000%		
50) Toluene-d8	8.653	98	109503	48.347	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.700%		
62) 4-Bromofluorobenzene	11.079	95	83017	47.937	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.880%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	2.995	59	7111	23.931	ug/l #	85
16) Acetone	2.392	43	3801m	5.285	ug/l	
30) Chloroform	5.092	83	63935	20.857	ug/l	98
47) Bromodichloromethane	7.824	83	13697	5.860	ug/l	99
60) Dibromochloromethane	9.525	129	2821	1.531	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033297.D
Acq On : 07 Dec 2022 21:28
Operator : JC/MD
Sample : N5911-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A3594TOG

Quant Time: Dec 08 01:44:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/08/2022
Supervised By :Mahesh Dadoda 12/08/2022

