

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033298.D
 Acq On : 07 Dec 2022 21:51
 Operator : JC/MD
 Sample : N5906-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Dec 08 01:44:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136672	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221390	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46311	52.818	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.640%	
35) Dibromofluoromethane	5.391	113	50218	50.431	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.860%	
50) Toluene-d8	8.653	98	103032	49.461	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.920%	
62) 4-Bromofluorobenzene	11.079	95	77065	48.384	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.760%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	2594	9.677	ug/l	99
16) Acetone	2.386	43	1508	2.324	ug/l	95
19) Methyl tert-butyl Ether	3.111	73	30984	7.219	ug/l	99
29) Tetrahydrofuran	5.025	42	3449	5.716	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033298.D
Acq On : 07 Dec 2022 21:51
Operator : JC/MD
Sample : N5906-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Dec 08 01:44:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

