

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
 Data File : VX033305.D
 Acq On : 08 Dec 2022 00:33
 Operator : JC/MD
 Sample : N5906-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24

Quant Time: Dec 08 01:46:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	157917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47547	46.933	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.860%
35) Dibromofluoromethane	5.391	113	53106	47.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.120%
50) Toluene-d8	8.653	98	112158	47.515	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.040%
62) 4-Bromofluorobenzene	11.079	95	82484	45.701	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.400%
Target Compounds						
					Qvalue	
30) Chloroform	5.099	83	1940	0.607	ug/l	99
64) Tetrachloroethene	9.281	164	489	0.331	ug/l	87
67) Ethyl Benzene	10.195	91	3644	0.477	ug/l #	87
78) n-propylbenzene	11.305	91	3526	0.445	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120722\
Data File : VX033305.D
Acq On : 08 Dec 2022 00:33
Operator : JC/MD
Sample : N5906-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24

Quant Time: Dec 08 01:46:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

