

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
 Data File : VX034357.D
 Acq On : 06 Mar 2023 18:09
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: Mar 07 00:42:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	279809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	471700	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	429278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	216596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	204145	54.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.340%
35) Dibromofluoromethane	5.385	113	164245	53.479	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.960%
50) Toluene-d8	8.647	98	610980	55.409	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.820%
62) 4-Bromofluorobenzene	11.079	95	250268	59.336	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.680%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	128376	53.095	ug/l	96
3) Chloromethane	1.300	50	179105	51.476	ug/l	99
4) Vinyl Chloride	1.374	62	179520	56.653	ug/l	99
5) Bromomethane	1.599	94	85473	74.367	ug/l	100
6) Chloroethane	1.678	64	102737	64.492	ug/l	98
7) Trichlorofluoromethane	1.886	101	268468	62.624	ug/l	98
8) Diethyl Ether	2.136	74	100280	58.062	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	155242	51.420	ug/l	95
10) Methyl Iodide	2.453	142	199812	54.357	ug/l	95
11) Tert butyl alcohol	2.959	59	222492	263.978	ug/l #	83
12) 1,1-Dichloroethene	2.319	96	153295	50.652	ug/l	96
13) Acrolein	2.233	56	115708	300.742	ug/l	99
14) Allyl chloride	2.666	41	313899	48.895	ug/l	91
15) Acrylonitrile	3.062	53	500213	272.049	ug/l	99
16) Acetone	2.380	43	468434	240.975	ug/l	92
17) Carbon Disulfide	2.514	76	404833	46.526	ug/l	98
18) Methyl Acetate	2.703	43	332554	55.288	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	574039	53.742	ug/l	97
20) Methylene Chloride	2.788	84	186225	53.168	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	169792	53.540	ug/l	94
22) Diisopropyl ether	3.757	45	637605	56.652	ug/l	96
23) Vinyl Acetate	3.721	43	2487828	275.897	ug/l	99
24) 1,1-Dichloroethane	3.611	63	330393	53.473	ug/l	98
25) 2-Butanone	4.556	43	747191	265.652	ug/l	99
26) 2,2-Dichloropropane	4.477	77	282564	50.483	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	203816	54.274	ug/l	95
28) Bromochloromethane	4.897	49	141668	51.398	ug/l	88
29) Tetrahydrofuran	5.001	42	471933	275.044	ug/l	98
30) Chloroform	5.092	83	332558	56.052	ug/l	99
31) Cyclohexane	5.470	56	295574	52.535	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	294500	55.373	ug/l	99
36) 1,1-Dichloropropene	5.696	75	244815	53.900	ug/l	98
37) Ethyl Acetate	4.714	43	282782	55.708	ug/l	99
38) Carbon Tetrachloride	5.678	117	250184	53.529	ug/l	99
39) Methylcyclohexane	7.379	83	295948	50.632	ug/l	97
40) Benzene	6.037	78	729649	54.208	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
 Data File : VX034357.D
 Acq On : 06 Mar 2023 18:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 07 00:42:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	157864	55.817	ug/l	99
42) 1,2-Dichloroethane	6.086	62	271367	57.323	ug/l	97
43) Isopropyl Acetate	6.342	43	490338	54.739	ug/l	99
44) Trichloroethene	7.123	130	186685	54.046	ug/l	91
45) 1,2-Dichloropropane	7.427	63	195453	54.547	ug/l	99
46) Dibromomethane	7.580	93	130421	55.305	ug/l #	85
47) Bromodichloromethane	7.824	83	259198	53.559	ug/l	94
48) Methyl methacrylate	7.690	41	232834	54.763	ug/l	97
49) 1,4-Dioxane	7.659	88	92270	1088.043	ug/l #	99
51) 4-Methyl-2-Pentanone	8.573	43	1453124	294.850	ug/l	99
52) Toluene	8.720	92	468172	57.403	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	304846	54.724	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	313748	52.610	ug/l	93
55) 1,1,2-Trichloroethane	9.153	97	190817	57.963	ug/l	98
56) Ethyl methacrylate	9.116	69	321662	57.758	ug/l	96
57) 1,3-Dichloropropane	9.311	76	328303	57.319	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	685535	284.163	ug/l	95
59) 2-Hexanone	9.427	43	1123677	299.175	ug/l	98
60) Dibromochloromethane	9.518	129	204452	56.206	ug/l	100
61) 1,2-Dibromoethane	9.610	107	202649	58.205	ug/l	98
64) Tetrachloroethene	9.275	164	157132	51.863	ug/l	93
65) Chlorobenzene	10.079	112	496387	55.473	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	188205	54.330	ug/l	98
67) Ethyl Benzene	10.195	91	908648	56.522	ug/l	98
68) m/p-Xylenes	10.299	106	688501	111.913	ug/l	93
69) o-Xylene	10.640	106	343435	56.074	ug/l	94
70) Styrene	10.652	104	568526	57.075	ug/l	95
71) Bromoform	10.799	173	150770	53.321	ug/l	99
73) Isopropylbenzene	10.963	105	888844	53.845	ug/l	99
74) N-amyl acetate	10.841	43	454146	54.762	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	301890	53.470	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	351765	64.542	ug/l	82
77) Bromobenzene	11.195	156	213665	53.786	ug/l	83
78) n-propylbenzene	11.305	91	1050604	54.724	ug/l	97
79) 2-Chlorotoluene	11.366	91	635748	54.005	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	748461	53.929	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	104053	49.045	ug/l	97
82) 4-Chlorotoluene	11.451	91	730927	54.514	ug/l	95
83) tert-Butylbenzene	11.713	119	755970	53.184	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	757099	53.953	ug/l	96
85) sec-Butylbenzene	11.890	105	924295	53.547	ug/l	97
86) p-Isopropyltoluene	12.012	119	766379	53.648	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	400424	53.748	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	404726	53.860	ug/l	96
89) n-Butylbenzene	12.335	91	681450	53.558	ug/l	98
90) Hexachloroethane	12.536	117	149639	50.099	ug/l	79
91) 1,2-Dichlorobenzene	12.335	146	391719	53.739	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	69682	51.756	ug/l	75
93) 1,2,4-Trichlorobenzene	13.585	180	243269	51.983	ug/l	98
94) Hexachlorobutadiene	13.725	225	94308	48.612	ug/l	98
95) Naphthalene	13.774	128	860177	54.563	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	243952	52.508	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
Data File : VX034357.D
Acq On : 06 Mar 2023 18:09
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Mar 07 00:42:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 279809
Ion Ratio Lower Upper
168 100
99 55.5 39.6 59.4



#2
Dichlorodifluoromethane
Concen: 53.095 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 85 Resp: 128376
Ion Ratio Lower Upper
85 100
87 33.7 15.8 47.4









#3
Chloromethane
Concen: 51.476 ug/l
RT: 1.300 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 179105
Ion Ratio Lower Upper
50 100
52 32.2 25.2 37.8



#4
Vinyl Chloride
Concen: 56.653 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 62 Resp: 179520
Ion Ratio Lower Upper
62 100
64 33.4 26.2 39.4









#5
Bromomethane
Concen: 74.367 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 85473
Ion Ratio Lower Upper
94 100
96 93.1 74.3 111.5



#6
Chloroethane
Concen: 64.492 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 64 Resp: 102737
Ion Ratio Lower Upper
64 100
66 31.2 26.1 39.1









#7
 Trichlorofluoromethane
 Concen: 62.624 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 268468
 Ion Ratio Lower Upper
 101 100
 103 64.6 50.6 75.8



#8
 Diethyl Ether
 Concen: 58.062 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Tgt Ion: 74 Resp: 100280
 Ion Ratio Lower Upper
 74 100
 45 98.0 51.5 154.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.420 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 155242
Ion Ratio Lower Upper
101 100
85 44.5 36.7 55.1
151 77.3 66.5 99.7



#10
Methyl Iodide
Concen: 54.357 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:142 Resp: 199812
Ion Ratio Lower Upper
142 100
127 42.6 37.3 55.9
141 14.3 11.5 17.3









#11
Tert butyl alcohol
Concen: 263.978 ug/l
RT: 2.959 min Scan# 306
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 222492
Ion Ratio Lower Upper
59 100
57 10.3 3.7 5.5#



#12
1,1-Dichloroethene
Concen: 50.652 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 96 Resp: 153295
Ion Ratio Lower Upper
96 100
61 160.4 134.2 201.4
98 62.7 50.2 75.2









#13
Acrolein
Concen: 300.742 ug/l
RT: 2.233 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 115708
Ion Ratio Lower Upper
56 100
55 69.3 56.3 84.5



#14
Allyl chloride
Concen: 48.895 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 41 Resp: 313899
Ion Ratio Lower Upper
41 100
39 63.8 59.0 88.4
76 32.5 24.6 37.0









#15
 Acrylonitrile
 Concen: 272.049 ug/l
 RT: 3.062 min Scan# 31
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 53 Resp: 500213
 Ion Ratio Lower Upper
 53 100
 52 82.9 65.7 98.5
 51 35.7 29.0 43.4



#16
 Acetone
 Concen: 240.975 ug/l
 RT: 2.380 min Scan# 212
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Tgt Ion: 43 Resp: 468434
 Ion Ratio Lower Upper
 43 100
 58 32.4 22.6 33.8









#17
Carbon Disulfide
Concen: 46.526 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 404833
Ion Ratio Lower Upper
76 100
78 9.4 7.0 10.6



#18
Methyl Acetate
Concen: 55.288 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 43 Resp: 332554
Ion Ratio Lower Upper
43 100
74 22.8 18.2 27.4









#19
Methyl tert-butyl Ether
Concen: 53.742 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 574039
Ion Ratio Lower Upper
73 100
57 23.0 17.4 26.0



#20
Methylene Chloride
Concen: 53.168 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 84 Resp: 186225
Ion Ratio Lower Upper
84 100
49 131.2 101.1 151.7
51 39.7 31.0 46.6
86 66.0 52.6 79.0









#21
trans-1,2-Dichloroethene
Concen: 53.540 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 169792
Ion Ratio Lower Upper
96 100
61 141.0 106.7 160.1
98 65.9 50.0 75.0



#22
Diisopropyl ether
Concen: 56.652 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 45 Resp: 637605
Ion Ratio Lower Upper
45 100
43 84.7 64.6 97.0
87 25.8 21.9 32.9
59 11.2 9.2 13.8









#23
 Vinyl Acetate
 Concen: 275.897 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 2487828
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 53.473 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Tgt Ion: 63 Resp: 330393
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.8 11.2
 100 4.2 2.4 7.2









#25
2-Butanone
Concen: 265.652 ug/l
RT: 4.556 min Scan# 50
Delta R.T. 0.012 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 747191
Ion Ratio Lower Upper
43 100
72 24.8 19.4 29.2



#26
2,2-Dichloropropane
Concen: 50.483 ug/l
RT: 4.477 min Scan# 556
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 77 Resp: 282564
Ion Ratio Lower Upper
77 100
97 22.5 11.9 35.5









#27
 cis-1,2-Dichloroethene
 Concen: 54.274 ug/l
 RT: 4.489 min Scan# 557
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 203816
 Ion Ratio Lower Upper
 96 100
 61 147.8 0.0 277.6
 98 64.4 0.0 128.0



#28
 Bromochloromethane
 Concen: 51.398 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. 0.006 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Tgt Ion: 49 Resp: 141668
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 4.6
 130 74.2 68.5 102.7





#29

Tetrahydrofuran

Concen: 275.044 ug/l

RT: 5.001 min Scan# 641

Delta R.T. 0.006 min

Lab File: VX034357.D


Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 42 Resp: 471933

Ion Ratio Lower Upper

42 100

72 42.7 34.8 52.2

71 39.1 32.2 48.2



#30

Chloroform

Concen: 56.052 ug/l

RT: 5.092 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion: 83 Resp: 332558

Ion Ratio Lower Upper

83 100


85 64.3 52.1 78.1








#31
Cyclohexane
Concen: 52.535 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC







Tgt Ion: 56 Resp: 295574
Ion Ratio Lower Upper
56 100
69 30.6 24.6 37.0
84 81.8 71.0 106.6



#32
1,1,1-Trichloroethane
Concen: 55.373 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 97 Resp: 294500
Ion Ratio Lower Upper
97 100
99 63.7 52.3 78.5
61 44.2 35.2 52.8





#33

1,2-Dichloroethane-d4

Concen: 54.173 ug/l

RT: 5.958 min Scan# 798

Delta R.T. 0.006 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 204145

Ion Ratio Lower Upper

65 100

67 52.8 0.0 104.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.006 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09




Tgt Ion: 114 Resp: 471700


Ion Ratio Lower Upper

114 100

63 21.0 0.0 37.4

88 16.2 0.0 30.6









#35
Dibromofluoromethane
Concen: 53.479 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 164245
Ion Ratio Lower Upper
113 100
111 104.4 83.0 124.6
192 19.8 19.0 28.4



#36
1,1-Dichloropropene
Concen: 53.900 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 75 Resp: 244815
Ion Ratio Lower Upper
75 100
110 35.2 18.8 56.4
77 31.2 25.1 37.7





#37

Ethyl Acetate

Concen: 55.708 ug/l

RT: 4.714 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 282782

Ion Ratio Lower Upper

43 100

61 12.9 11.1 16.7

70 10.5 8.5 12.7



#38

Carbon Tetrachloride

Concen: 53.529 ug/l

RT: 5.678 min Scan# 753

Delta R.T. 0.006 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09




Tgt Ion: 117 Resp: 250184


Ion Ratio Lower Upper

117 100

119 97.0 78.0 117.0

121 32.4 24.4 36.6









#39
Methylcyclohexane
Concen: 50.632 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 295948
Ion Ratio Lower Upper
83 100
55 80.0 61.8 92.6
98 47.8 39.3 58.9



#40
Benzene
Concen: 54.208 ug/l
RT: 6.037 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 78 Resp: 729649
Ion Ratio Lower Upper
78 100
77 24.0 19.0 28.4









#41
Methacrylonitrile
Concen: 55.817 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	157864
Ion	Ratio	Lower	Upper
41	100		
39	55.3	43.8	65.6
67	69.3	54.8	82.2
52	29.0	23.0	34.6



#42
1,2-Dichloroethane
Concen: 57.323 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:	62	Resp:	271367
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	20.2









#43
Isopropyl Acetate
Concen: 54.739 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 490338
Ion Ratio Lower Upper
43 100
61 19.5 15.7 23.5
87 12.6 10.6 15.8



#44
Trichloroethene
Concen: 54.046 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:130 Resp: 186685
Ion Ratio Lower Upper
130 100
95 99.9 0.0 182.6









#45
1,2-Dichloropropane
Concen: 54.547 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 63 Resp: 195453
Ion Ratio Lower Upper
63 100
65 29.4 24.2 36.2



#46
Dibromomethane
Concen: 55.305 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 93 Resp: 130421
Ion Ratio Lower Upper
93 100
95 81.9 67.6 101.4
174 96.1 97.4 146.0#









#47
Bromodichloromethane
Concen: 53.559 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 259198
Ion Ratio Lower Upper
83 100
85 67.0 49.8 74.6
127 8.2 7.0 10.4



#48
Methyl methacrylate
Concen: 54.763 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 41 Resp: 232834
Ion Ratio Lower Upper
41 100
69 79.0 63.1 94.7
39 58.7 43.2 64.8





#49

1,4-Dioxane

Concen: 1088.043 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 92270

Ion Ratio Lower Upper

88 100

43 33.6 0.0 0.0#

58 71.2 56.5 84.7



#50

Toluene-d8

Concen: 55.409 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion: 98 Resp: 610980

Ion Ratio Lower Upper

98 100

100 65.4 53.6 80.4









#51
4-Methyl-2-Pentanone
Concen: 294.850 ug/l
RT: 8.573 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 1453124
Ion Ratio Lower Upper
43 100
58 38.4 31.2 46.8



#52
Toluene
Concen: 57.403 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 92 Resp: 468172
Ion Ratio Lower Upper
92 100
91 171.2 135.5 203.3









#53
t-1,3-Dichloropropene
Concen: 54.724 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 304846
Ion Ratio Lower Upper
75 100
77 31.7 25.5 38.3



#54
cis-1,3-Dichloropropene
Concen: 52.610 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 75 Resp: 313748
Ion Ratio Lower Upper
75 100
77 32.7 25.0 37.6
39 54.8 38.6 58.0









#55
 1,1,2-Trichloroethane
 Concen: 57.963 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.000 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 97 Resp: 190817




Ion	Ratio	Lower	Upper
97	100		
83	85.2	67.4	101.0
85	56.2	43.2	64.8
99	64.8	50.2	75.2




#56
 Ethyl methacrylate
 Concen: 57.758 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Tgt Ion: 69 Resp: 321662

Ion	Ratio	Lower	Upper
69	100		
41	70.9	55.1	82.7
39	40.5	28.5	42.7





#57

1,3-Dichloropropane

Concen: 57.319 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.006 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 328303

Ion Ratio Lower Upper

76 100

78 32.5 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 284.163 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion: 63 Resp: 685535

Ion Ratio Lower Upper

63 100

106 27.8 24.5 36.7





#59

2-Hexanone

Concen: 299.175 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 1123677

Ion Ratio Lower Upper

43 100

58 53.2 27.4 82.0



#60

Dibromochloromethane

Concen: 56.206 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion: 129 Resp: 204452

Ion Ratio Lower Upper

129 100

127 77.4 38.8 116.3





#61

1,2-Dibromoethane

Concen: 58.205 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:107 Resp: 202649

Ion Ratio Lower Upper

107 100

109 94.5 74.2 111.2



#62

4-Bromofluorobenzene

Concen: 59.336 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09




Tgt Ion: 95 Resp: 250268


Ion Ratio Lower Upper

95 100

174 73.0 0.0 184.4

176 70.2 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 429278
Ion Ratio Lower Upper
117 100
82 57.4 40.6 60.8
119 32.5 25.6 38.4



#64
Tetrachloroethene
Concen: 51.863 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:164 Resp: 157132
Ion Ratio Lower Upper
164 100
166 127.9 100.2 150.2
129 96.1 69.2 103.8
131 92.4 67.2 100.8









#65
Chlorobenzene
Concen: 55.473 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 496387
Ion Ratio Lower Upper
112 100
114 31.4 26.2 39.4



#66
1,1,1,2-Tetrachloroethane
Concen: 54.330 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:131 Resp: 188205
Ion Ratio Lower Upper
131 100
133 95.2 48.4 145.0
119 65.3 31.9 95.7









#67
Ethyl Benzene
Concen: 56.522 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 908648
Ion Ratio Lower Upper
91 100
106 31.5 25.8 38.8



#68
m/p-Xylenes
Concen: 111.913 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 106 Resp: 688501
Ion Ratio Lower Upper
106 100
91 205.4 156.4 234.6









#69
o-Xylene
Concen: 56.074 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 343435
Ion Ratio Lower Upper
106 100
91 216.0 103.1 309.3



#70
Styrene
Concen: 57.075 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:104 Resp: 568526
Ion Ratio Lower Upper
104 100
78 53.6 37.5 56.3
103 54.9 43.3 64.9









#71
Bromoform
Concen: 53.321 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 150770
Ion Ratio Lower Upper
173 100
175 49.0 24.1 72.3
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:152 Resp: 216596
Ion Ratio Lower Upper
152 100
115 89.2 38.7 116.1
150 176.6 0.0 348.2





#73

Isopropylbenzene

Concen: 53.845 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 888844

Ion Ratio Lower Upper

105 100

120 26.8 13.6 40.8



#74

N-ethyl acetate

Concen: 54.762 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Tgt Ion: 43 Resp: 454146




Ion Ratio Lower Upper


43 100

70 40.4 34.8 52.2

55 22.4 19.4 29.0

61 24.1 19.4 29.2









#75
 1,1,2,2-Tetrachloroethane
 Concen: 53.470 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 301890
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.5 16.7
 85 64.5 32.3 96.9



#76
 1,2,3-Trichloropropane
 Concen: 64.542 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. 0.000 min
 Lab File: VX034357.D
 Acq: 06 Mar 2023 18:09

Tgt Ion: 75 Resp: 351765
 Ion Ratio Lower Upper
 75 100
 77 32.0 21.8 65.4





#77

Bromobenzene

Concen: 53.786 ug/l

RT: 11.195 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 213665

Ion Ratio Lower Upper

156 100

77 171.7 68.1 204.3

158 96.0 48.0 144.2



#78

n-propylbenzene

Concen: 54.724 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion: 91 Resp: 1050604

Ion Ratio Lower Upper

91 100

120 22.9 12.3 36.8









#79
2-Chlorotoluene
Concen: 54.005 ug/l
RT: 11.366 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 635748
Ion Ratio Lower Upper
91 100
126 33.0 18.1 54.1



#80
1,3,5-Trimethylbenzene
Concen: 53.929 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 105 Resp: 748461
Ion Ratio Lower Upper
105 100
120 48.3 25.6 76.8









#81
trans-1,4-Dichloro-2-butene
Concen: 49.045 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 104053
Ion Ratio Lower Upper
75 100
53 100.2 83.0 124.6
89 45.8 37.7 56.5



#82
4-Chlorotoluene
Concen: 54.514 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 91 Resp: 730927
Ion Ratio Lower Upper
91 100
126 28.8 15.8 47.4





#83

tert-Butylbenzene

Concen: 53.184 ug/l

RT: 11.713 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:119 Resp: 755970

Ion Ratio Lower Upper

119 100

91 59.6 26.9 80.7

134 23.0 12.3 36.8



#84

1,2,4-Trimethylbenzene

Concen: 53.953 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion:105 Resp: 757099

Ion Ratio Lower Upper

105 100

120 44.4 23.4 70.2





#85

sec-Butylbenzene

Concen: 53.547 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 924295

Ion Ratio Lower Upper

105 100

134 20.3 10.8 32.3



#86

p-Isopropyltoluene

Concen: 53.648 ug/l

RT: 12.012 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09




Tgt Ion:119 Resp: 766379


Ion Ratio Lower Upper

119 100

134 26.8 13.6 40.8

91 25.7 11.0 33.0









#87
1,3-Dichlorobenzene
Concen: 53.748 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 400424
Ion Ratio Lower Upper
146 100
111 42.0 18.6 55.6
148 63.8 31.9 95.5



#88
1,4-Dichlorobenzene
Concen: 53.860 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:146 Resp: 404726
Ion Ratio Lower Upper
146 100
111 41.0 17.8 53.4
148 63.8 32.1 96.3





#89

n-Butylbenzene

Concen: 53.558 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 681450

Ion Ratio Lower Upper

91 100

92 54.7 27.5 82.5

134 25.8 14.3 42.9



#90

Hexachloroethane

Concen: 50.099 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX034357.D




Acq: 06 Mar 2023 18:09


Tgt Ion: 117 Resp: 149639

Ion Ratio Lower Upper

117 100

201 70.1 45.1 135.3









#91
1,2-Dichlorobenzene
Concen: 53.739 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 391719
Ion Ratio Lower Upper
146 100
111 43.9 19.3 57.8
148 63.9 32.3 96.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.756 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion: 75 Resp: 69682
Ion Ratio Lower Upper
75 100
155 80.7 51.7 155.1
157 103.5 67.5 202.3









#93
1,2,4-Trichlorobenzene
Concen: 51.983 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.006 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:180 Resp: 243269
Ion Ratio Lower Upper
180 100
182 95.7 47.9 143.6
145 32.1 14.1 42.2



#94
Hexachlorobutadiene
Concen: 48.612 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX034357.D
Acq: 06 Mar 2023 18:09

Tgt Ion:225 Resp: 94308
Ion Ratio Lower Upper
225 100
223 61.5 31.3 93.9
227 63.3 32.3 96.9





#95

Naphthalene

Concen: 54.563 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 860177

Ion Ratio Lower Upper

128 100

127 13.0 10.3 15.5

129 11.0 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 52.508 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX034357.D

Acq: 06 Mar 2023 18:09

Tgt Ion:180 Resp: 243952

Ion Ratio Lower Upper

180 100

182 96.3 47.6 142.9

145 34.3 15.1 45.3

